

YOUR REF. NO.: -
OF DATE:
OUR REF. NO.: CMI-6015/2019/025

HANDLED BY: Jindřich Bílek
PHONE: +420 545 555 717
E-MAIL: jbilek2@cmi.cz

DATE: 2019-09-13

Start Italiana s.r.l.
Via Pola 6
20813 Bovisio Masciago (MB)

D11 compliance declaration

Dear Mr. Caligaris

Regarding your question about compliance of level gauges certified according OIML R85 with document OIML D11 can be declared following:

OIML D11 is so-called internal document of OIML. It is very general document focused on requirements for measuring instruments from the point of view of environmental conditions.

As it is stated in its scope, this Document shall be taken into account by OIML Technical Committees, Subcommittees and Project Groups when establishing performance requirements and tests for a specific category of measuring instruments concerning their sensitivity to influence quantities.

This means that it is a base for creating the "R" documents (recommendations) and is not directly intended for certification of measuring instruments.

Document OIML D11 is mentioned in bibliography of document OIML R85. This means that level gauges certified according OIML R85 is in compliance with relevant parts of document OIML D11.

Best regards



RNDr. Pavel Klenovský
General Director

Český metrologický institut

Okružní 31
638 00 Brno
-19-